
Certificate of Irradiation

Product	Nunc 10 µl Inoculation Loop, Blue	Date 2025-06-19
		Page 1
Item Number	251586	
Lot Number	185590	
Specified Dose Range	22.4 - 55.5 kGy	
SAL	10E-6	
Quantity	168 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1299671A	DK00491262	2025-06-18	48.5
EC__1299657A	DK00491261	2025-06-18	49.3

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature