
Certificate of Irradiation

Product	Nunc Tubes, Nunclon, flat, w. screw cap, 3 ml	Date 2025-10-28
		Page 1
Item Number	156758	
Lot Number	184562	
Specified Dose Range	21.5 - 38.5 kGy	
SAL	10E-6	
Quantity	120 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1302470B	DK00491545	2025-10-25	26.3
EC__1302470A	DK00491546	2025-10-25	26.1
EC__1302498A	DK00491547	2025-10-25	26.2

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature