

Certificate of Irradiation

Product	Nunc Tubes, Nunclon, flat, w. screw cap, 3 ml	Date 2025-10-28
		Page 1
Item Number	156758	
Lot Number	184574	
Specified Dose Range	21.5 - 38.5 kGy	
SAL	10E-6	
Quantity	160 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC_1303841A	DK00491514	2025-10-16	25.5
EC_1303827A	DK00491515	2025-10-16	25.9
EC_1303839A	DK00491517	2025-10-16	25.6
EC_1303771A	DK00491516	2025-10-16	25.6

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack

Senior Quality Assurance Manager

Please note that this is an electronically generated signature