

Certificate of Irradiation

Product	Nunc Dish Nunclon 245 x 245 x 25	Date 2025-09-02
		Page 1
Item Number	166508	
Lot Number	186489	
Specified Dose Range	22.2 - 40.3 kGy	
SAL	10E-6	
Quantity	280 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1724342A	DK00474340	2025-08-08	28.7
EC__1300868A	DK00474341	2025-08-23	25.6
EC__1300868A	DK00474341	2025-08-23	25.6
EC__1300896A	DK00474342	2025-08-23	26.5
EC__1300896A	DK00474342	2025-08-23	26.5
EC__1300322A	DK00474343	2025-08-23	27.2
EC__1300322A	DK00474343	2025-08-23	27.2
EC__1300350A	DK00474344	2025-08-23	26.9
EC__1300350A	DK00474344	2025-08-23	26.9
EC__1300756A	DK00474345	2025-08-23	26.7
EC__1300756A	DK00474345	2025-08-23	26.7
EC__1300364A	DK00474346	2025-08-23	27.9
EC__1300364A	DK00474346	2025-08-23	27.9

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature