
Certificate of Irradiation

Product	Nunc EasYFlasks, w. Filter	Date 2025-07-09
		Page 1
Item Number	156800	
Lot Number	186222	
Specified Dose Range	16.0 - 45.7 kGy	
SAL	10E-6	
Quantity	260 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1723992A	DK00497007	2025-07-08	17.9
EA__1724035B	DK00481333	2025-07-08	18.2
EA__1723966A	DK00481334	2025-07-08	17.9
EA__1724069B	DK00481335	2025-07-08	17.9
EA__1724036B	DK00481336	2025-07-08	18.2
EA__1724052B	DK00497003	2025-07-08	17.9
EA__1724100B	DK00497004	2025-07-08	17.9
EA__1724116B	DK00497005	2025-07-08	18.7
EA__1724068B	DK00497006	2025-07-08	17.8
EA__1724012A	DK00497008	2025-07-08	17.8
EA__1724008A	DK00497009	2025-07-08	17.9
EA__1723892B	DK00497010	2025-07-08	17.8
EA__1723948B	DK00497011	2025-07-09	18.2

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature