

Certificate of Irradiation

Product	Nunc 6 Well Multidishes, Cell Culture Treated	Date 2025-06-19
		Page 1
Item Number	140675	
Lot Number	186393	
Specified Dose Range	15.8 - 45.0 kGy	
SAL	10E-3	
Quantity	674 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300738B	DK00484560	2025-06-05	18.9
EC__1300536A	DK00484561	2025-06-05	19.1
EC__1300467A	DK00484562	2025-06-05	19.2
EC__1300578B	DK00484563	2025-06-05	19.3
EC__1300385B	DK00484564	2025-06-05	19.3
EC__1300442A	DK00484565	2025-06-05	19.3
EC__1300536B	DK00484566	2025-06-05	19.3
EC__1300537B	DK00484567	2025-06-05	19.3
EC__1300792B	DK00484568	2025-06-05	19.3
EC__1300840B	DK00484569	2025-06-05	18.9
EC__1300808B	DK00484570	2025-06-05	19.3
EC__1300564B	DK00484571	2025-06-05	19.3
EC__1299704B	DK00484572	2025-06-18	19.4
EC__1299721B	DK00484574	2025-06-18	19.3
EC__1301007B	DK00484573	2025-06-18	19.2

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature