

Certificate of Irradiation

Product	Nunc IVF Petri Dishes, Non Treated	Date 2025-04-08
		Page 1
Item Number	150255	
Lot Number	184962	
Specified Dose Range	24.2 - 37.9 kGy	
SAL	10E-6	
Quantity	1040 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1719146B	DK00397533	2025-04-07	32.2
EA__1719147B	DK00397532	2025-04-07	31.9
EA__1719162B	DK00397534	2025-04-07	31.3
EA__1719145A	DK00397535	2025-04-07	31.0
EA__1719161A	DK00397536	2025-04-07	31.7
EA__1719144B	DK00397537	2025-04-07	31.7
EA__1719129A	DK00397538	2025-04-07	31.4
EA__1719131A	DK00397539	2025-04-07	30.9
EA__1719113A	DK00397540	2025-04-07	30.7
EA__1719121B	DK00397541	2025-04-07	31.6
EA__1719105B	DK00397542	2025-04-07	31.4
EA__1719129B	DK00397544	2025-04-07	30.8
EA__1719113B	DK00397545	2025-04-07	33.7

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature