

## Certificate of Irradiation

<b>Product</b>	Nunc IVF Petri Dishes, Non Treated	<b>Date</b> 2025-05-13
		<b>Page</b> 1
<b>Item Number</b>	150255	
<b>Lot Number</b>	185375	
<b>Specified Dose Range</b>	24.2 - 37.9 kGy	
<b>SAL</b>	10E-6	
<b>Quantity</b>	1040 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1299640A	DK00397550	2025-05-12	34.4
EC__1299612A	DK00397551	2025-05-12	30.0
EC__1301095B	DK00397552	2025-05-12	30.8
EC__1299654A	DK00397553	2025-05-12	30.5
EC__1299597B	DK00397554	2025-05-12	30.4
EC__1300634B	DK00397555	2025-05-12	30.3
EC__1300662B	DK00397556	2025-05-12	30.3
EC__1301129B	DK00397557	2025-05-12	30.2
EC__1300718B	DK00397558	2025-05-12	30.4
EC__1299567B	DK00397559	2025-05-12	31.1
EC__1299632B	DK00397560	2025-05-12	31.1
EC__1299589B	DK00397561	2025-05-12	31.0
EC__1301133B	DK00397562	2025-05-12	30.5

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager

Please note that this is an electronically generated signature