

Certificate of Irradiation

Product	Nunc IVF Petri Dishes, Non Treated	Date 2025-06-11
		Page 1
Item Number	150255	
Lot Number	185608	
Specified Dose Range	24.2 - 37.9 kGy	
SAL	10E-6	
Quantity	1043 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300412A	DK00397574	2025-06-10	31.1
EC__1300817B	DK00397584	2025-06-10	31.8
EC__1300370A	DK00397575	2025-06-10	29.8
EC__1300440A	DK00397576	2025-06-10	29.8
EC__1300859B	DK00397578	2025-06-10	34.5
EC__1300684A	DK00397579	2025-06-10	31.3
EC__1300391A	DK00397580	2025-06-10	31.9
EC__1300357A	DK00397581	2025-06-10	30.4
EC__1300845B	DK00397582	2025-06-10	30.9
EC__1300831B	DK00397583	2025-06-10	31.7
EC__1300872B	DK00397588	2025-06-10	32.3
EC__1300802A	DK00397585	2025-06-10	32.1
EC__1300830B	DK00397586	2025-06-10	31.8
EC__1300746B	DK00397587	2025-06-10	30.6

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature