

Certificate of Irradiation

Product	Nunc Flasks with Filter Caps	Date 2025-05-04
		Page 1
Item Number	178905	
Lot Number	185630	
Specified Dose Range	17.1 - 47.8 kGy	
SAL	10E-6	
Quantity	767 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1718916B	DK00424224	2025-04-20	35.4
EA__1718900B	DK00424223	2025-04-20	34.5
EA__1718912B	DK00424222	2025-04-20	33.2
EA__1718901B	DK00424221	2025-04-20	33.5
EA__1718910A	DK00424220	2025-04-20	31.8
EA__1717706A	DK00424219	2025-04-20	29.3
EA__1718926A	DK00424218	2025-04-20	28.8
EA__1718922A	DK00424217	2025-04-20	28.3
EA__1718925B	DK00424216	2025-04-20	29.6
EA__1717711A	DK00424215	2025-04-20	29.4
EA__1717722B	DK00424227	2025-04-20	34.7
EA__1718899A	DK00424226	2025-04-20	33.5
EA__1718915A	DK00424225	2025-04-20	33.9
EA__1724237A	DK00424229	2025-05-03	34.1
EA__1724278A	DK00424230	2025-05-03	34.7
EA__1724587A	DK00424228	2025-05-03	32.8

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature