
Certificate of Irradiation

Product	Nunc Flasks with Filter Caps	Date 2025-05-12
		Page 1
Item Number	178905	
Lot Number	185882	
Specified Dose Range	17.1 - 47.8 kGy	
SAL	10E-6	
Quantity	384 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300730B	DK00424242	2025-05-11	27.2
EC__1300649A	DK00424243	2025-05-11	27.2
EC__1301099B	DK00424245	2025-05-12	27.5
EC__1300971B	DK00424244	2025-05-12	28.0
EC__1300999B	DK00424246	2025-05-12	26.8
EC__1299649B	DK00424247	2025-05-12	27.1
EC__1301071B	DK00424248	2025-05-12	27.2
EC__1301118B	DK00424249	2025-05-12	26.8

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature