

Certificate of Irradiation

Product	Nunc F96 MicroWell Polystyrene Plate, White	Date 2025-08-19
		Page 1
Item Number	136102	
Lot Number	185579	
Specified Dose Range	16.4 - 45.9 kGy	
SAL	10E-3	
Quantity	160 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300134A	DK00511792	2025-08-16	18.8
EC__1300156A	DK00511793	2025-08-16	18.9
EC__1300005B	DK00511794	2025-08-16	19.2
EC__1300106A	DK00511795	2025-08-16	19.2
EC__1300026B	DK00511796	2025-08-16	19.0
EC__1300069A	DK00511797	2025-08-16	19.2
EC__1300057A	DK00511798	2025-08-16	18.6
EC__1300153B	DK00511799	2025-08-16	18.4

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature