

Certificate of Irradiation

Product	Nunc F96 MicroWell Plates Polystyrene, Clear	Date 2025-09-18
		Page 1
Item Number	156545	
Lot Number	186959	
Specified Dose Range	16.4 - 45.9 kGy	
SAL	10E-3	
Quantity	144 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC_1302663B	DK00517489	2025-09-05	18.5
EC_1302532A	DK00517496	2025-09-05	19.8
EC_1302560A	DK00517495	2025-09-05	19.8
EC_1302673A	DK00517494	2025-09-05	19.6
EC_1302588A	DK00517493	2025-09-05	19.3
EC_1302602A	DK00517492	2025-09-05	19.5
EC_1302686A	DK00517491	2025-09-05	19.6
EC_1302649B	DK00517490	2025-09-05	19.7

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature