

## Certificate of Irradiation

**Product** Nunc 48 Well Multidishes

Date 2025-09-18  
Page 1

**Item Number** 152640

**Lot Number** 186220

**Specified Dose Range** 14.8 - 45.5 kGy

**SAL** 10E-3

**Quantity** 64 boxes

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC_1301543B	DK00471630	2025-09-13	17.9

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager

Please note that this is an electronically generated signature