

Certificate of Irradiation

Product	Nunc F96 MicroWell Polystyrene Plate, Black	Date 2025-05-25
		Page 1
Item Number	137103	
Lot Number	185429	
Specified Dose Range	16.4 - 45.9 kGy	
SAL	10E-3	
Quantity	160 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300284B	DK00511727	2025-05-09	18.2
EC__1300217A	DK00511726	2025-05-09	18.6
EC__1300909A	DK00511730	2025-05-09	18.7
EC__1301089B	DK00511729	2025-05-09	18.2
EC__1300291A	DK00511728	2025-05-09	18.2
EC__1300195B	DK00511733	2025-05-24	18.9
EC__1300181B	DK00511732	2025-05-24	18.6
EC__1299569B	DK00511731	2025-05-24	18.8

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature