

---

## Certificate of Irradiation

<b>Product</b>	Nunc F96 Microwell plate, Non-Treated, White	<b>Date</b> 2025-06-28
		<b>Page</b> 1
<b>Item Number</b>	236105	
<b>Lot Number</b>	185632	
<b>Specified Dose Range</b>	17.1 - 45.9 kGy	
<b>SAL</b>	10E-3	
<b>Quantity</b>	73 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300439A	DK00511789	2025-06-05	19.0
EA__1723950A	DK00511790	2025-06-27	20.5

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager

Please note that this is an electronically generated signature