

Certificate of Irradiation

Product	Nunc EasYFlasks 75cm ² w. Filter	Date 2025-06-19
		Page 1
Item Number	156499	
Lot Number	185449	
Specified Dose Range	16.0 - 45.7 kGy	
SAL	10E-6	
Quantity	2040 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300283A	DK00474919	2025-05-23	24.8
EC__1300297A	DK00474920	2025-05-23	24.5
EC__1298757A	DK00474921	2025-05-23	25.2
EC__1300268B	DK00474922	2025-05-23	24.8
EC__1300310B	DK00474923	2025-05-23	25.1
EC__1300232B	DK00474924	2025-05-23	24.9
EC__1300274B	DK00474925	2025-05-23	25.1
EC__1300246B	DK00474926	2025-05-23	24.9
EC__1300176B	DK00474927	2025-05-23	24.6
EC__1300204B	DK00474928	2025-05-23	23.3
EC__1300190B	DK00474929	2025-05-23	20.6
EC__1299614B	DK00474930	2025-05-23	20.4
EC__1301123A	DK00474931	2025-05-23	19.7
EC__1301112B	DK00474933	2025-05-23	18.6
EC__1299642B	DK00474932	2025-05-23	19.2
EC__1299893B	DK00474938	2025-05-25	26.2
EC__1299879B	DK00474937	2025-05-25	25.4
EC__1299907B	DK00474939	2025-05-25	28.0
EC__1299865B	DK00474935	2025-05-25	18.9
EC__1299851B	DK00474934	2025-05-25	18.2
EC__1298752B	DK00474936	2025-05-25	19.2
EC__1299921B	DK00474940	2025-05-25	29.3
EC__1299935B	DK00474941	2025-05-25	29.9
EC__1299977B	DK00474942	2025-05-25	30.3
EC__1301088A	DK00474943	2025-05-25	29.1
EC__1296262A	DK00474944	2025-05-25	28.9
EC__1300016B	DK00474945	2025-05-25	29.6
EC__1300030B	DK00474946	2025-05-25	28.5
EC__1300044B	DK00474947	2025-05-25	28.9
EC__1300058B	DK00474948	2025-05-25	30.2
EC__1299836B	DK00474949	2025-05-27	29.3
EC__1298755A	DK00474950	2025-05-27	29.4
EC__1300207B	DK00474951	2025-05-27	29.6

Certificate of Irradiation

Product	Nunc EasYFlasks 75cm ² w. Filter	Date 2025-06-19
		Page 2
Item Number	156499	
Lot Number	185449	
Specified Dose Range	16.0 - 45.7 kGy	
SAL	10E-6	
Quantity	2040 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1299933A	DK00474952	2025-05-27	30.2
EC__1299961A	DK00474953	2025-05-27	28.2
EC__1299711A	DK00474954	2025-05-27	27.3
EC__1299725A	DK00474955	2025-05-27	26.0
EC__1299697A	DK00474956	2025-05-27	25.7
EC__1299683A	DK00474957	2025-05-27	24.7
EC__1299669A	DK00474958	2025-05-27	24.5
EC__1299820B	DK00474959	2025-05-27	24.5
EC__1300138B	DK00474963	2025-05-27	24.5
EC__1300078A	DK00474962	2025-05-27	24.5
EC__1300068B	DK00474961	2025-05-27	25.1
EC__1300092A	DK00474960	2025-05-27	24.7
EC__1300758B	DK00474964	2025-05-29	24.3
EC__1300772B	DK00474965	2025-05-29	24.9
EC__1300437B	DK00474966	2025-05-29	24.9
EC__1300373A	DK00474967	2025-05-29	24.9
EC__1300316A	DK00474968	2025-05-29	25.3
EC__1300316B	DK00474969	2025-05-29	25.5
EC__1300887B	DK00474970	2025-05-29	26.2
EC__1300413A	DK00474971	2025-05-29	25.2
EC__1300395B	DK00474972	2025-05-29	25.7
EC__1300325B	DK00474973	2025-05-29	25.5
EC__1300409B	DK00474978	2025-05-29	25.1
EC__1300773A	DK00474974	2025-05-29	24.1
EC__1300787A	DK00474975	2025-05-29	24.8
EC__1300382A	DK00474976	2025-05-29	25.2
EC__1300344B	DK00474977	2025-05-29	25.2
EC__1300850B	DK00474979	2025-06-04	31.2
EC__1300836B	DK00474980	2025-06-04	30.1
EC__1300822B	DK00474981	2025-06-04	29.8
EC__1300794B	DK00474982	2025-06-04	29.9
EC__1300780B	DK00474983	2025-06-04	29.5
EC__1300752B	DK00474984	2025-06-04	28.8

Certificate of Irradiation

Product	Nunc EasYFlasks 75cm ² w. Filter	Date 2025-06-19
		Page 3
Item Number	156499	
Lot Number	185449	
Specified Dose Range	16.0 - 45.7 kGy	
SAL	10E-6	
Quantity	2040 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300878B	DK00474985	2025-06-04	29.2
EC__1300338B	DK00474986	2025-06-04	29.0
EC__1300785B	DK00474987	2025-06-04	29.6
EC__1300771B	DK00474988	2025-06-04	29.2
EC__1300789A	DK00474989	2025-06-04	30.2
EC__1300341A	DK00474990	2025-06-04	25.5
EC__1300380B	DK00474991	2025-06-04	22.9
EC__1300408B	DK00474992	2025-06-04	21.5
EC__1300394B	DK00474993	2025-06-04	21.0
EC__1300366B	DK00474994	2025-06-04	19.8
EC__1300867B	DK00474995	2025-06-04	19.8
EC__1300339B	DK00474996	2025-06-04	19.5
EC__1300422B	DK00474997	2025-06-04	19.6
EC__1300595A	DK00474998	2025-06-04	20.0
EC__1300848B	DK00475006	2025-06-05	25.8
EC__1300416B	DK00474999	2025-06-05	19.7
EC__1300402B	DK00475001	2025-06-05	19.4
EC__1300318B	DK00475000	2025-06-05	19.8
EC__1300456A	DK00475002	2025-06-05	24.1
EC__1300581A	DK00475003	2025-06-05	24.3
EC__1300369A	DK00475004	2025-06-05	24.1
EC__1300862B	DK00475005	2025-06-05	25.3
EC__1300355A	DK00475019	2025-06-08	31.1
EC__1300401B	DK00475018	2025-06-08	31.1
EC__1300870A	DK00475017	2025-06-08	30.7
EC__1300884A	DK00475016	2025-06-08	30.3
EC__1300759B	DK00475015	2025-06-08	30.8
EC__1300898A	DK00475014	2025-06-08	30.3
EC__1300745B	DK00475013	2025-06-08	30.5
EC__1300828A	DK00475012	2025-06-08	31.4
EC__1300383A	DK00475011	2025-06-08	31.0
EC__1300744A	DK00475010	2025-06-08	31.1
EC__1301037B	DK00475020	2025-06-11	19.0

Certificate of Irradiation

Product	Nunc EasYFlasks 75cm ² w. Filter	Date 2025-06-19
		Page 4
Item Number	156499	
Lot Number	185449	
Specified Dose Range	16.0 - 45.7 kGy	
SAL	10E-6	
Quantity	2040 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
---------------------	----------	---------------------	----------

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature