

Certificate of Irradiation

Product	Nunc EasYFlasks 75cm ² w. Filter	Date 2025-06-19
		Page 1
Item Number	156499	
Lot Number	185537	
Specified Dose Range	16.0 - 45.7 kGy	
SAL	10E-6	
Quantity	798 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300503A	DK00481212	2025-05-31	25.2
EC__1300332A	DK00481211	2025-05-31	25.3
EC__1300318A	DK00481210	2025-05-31	25.4
EC__1300412B	DK00481217	2025-05-31	25.3
EC__1300489A	DK00481216	2025-05-31	24.8
EC__1300475A	DK00481215	2025-05-31	25.4
EC__1300574A	DK00481214	2025-05-31	24.6
EC__1300517A	DK00481213	2025-05-31	25.0
EC__1300555A	DK00481218	2025-05-31	29.2
EC__1300569A	DK00481219	2025-05-31	24.1
EC__1300583A	DK00481220	2025-05-31	24.1
EC__1300541A	DK00481221	2025-05-31	26.3
EC__1300461A	DK00481222	2025-05-31	25.7
EC__1300527A	DK00481223	2025-05-31	27.5
EC__1300513A	DK00481224	2025-05-31	28.9
EC__1300582A	DK00481225	2025-05-31	29.4
EC__1300568A	DK00481226	2025-05-31	29.8
EC__1300554A	DK00481227	2025-05-31	29.9
EC__1300540A	DK00481228	2025-05-31	31.0
EC__1300499A	DK00481229	2025-05-31	29.5
EC__1300485A	DK00481230	2025-05-31	29.8
EC__1300371B	DK00481231	2025-05-31	30.5
EC__1300357B	DK00481232	2025-05-31	30.2
EC__1300398B	DK00481233	2025-05-31	30.3
EC__1300384B	DK00481234	2025-05-31	30.8
EC__1300370B	DK00481235	2025-05-31	30.4
EC__1300844B	DK00481237	2025-06-10	30.2
EC__1300858B	DK00481236	2025-06-10	29.6
EC__1300346B	DK00481238	2025-06-11	30.3
EC__1300803B	DK00481239	2025-06-11	27.3
EC__1300789B	DK00481240	2025-06-11	21.8
EC__1300747B	DK00481241	2025-06-11	20.4
EC__1300900B	DK00481242	2025-06-11	19.2

Certificate of Irradiation

Product	Nunc EasYFlasks 75cm ² w. Filter	Date 2025-06-19
		Page 2
Item Number	156499	
Lot Number	185537	
Specified Dose Range	16.0 - 45.7 kGy	
SAL	10E-6	
Quantity	798 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300886B	DK00481243	2025-06-11	18.2
EC__1300396A	DK00481244	2025-06-11	17.8
EC__1300775B	DK00481245	2025-06-11	17.9
EC__1301031B	DK00481246	2025-06-11	19.9
EC__1301017B	DK00481247	2025-06-11	24.4
EC__1301003B	DK00481248	2025-06-11	25.8
EC__1301060B	DK00481249	2025-06-11	25.4

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature