

Certificate of Irradiation

Product	Nunc EasYFlasks 75cm ² w. Filter	Date 2025-06-19
		Page 1
Item Number	156499	
Lot Number	185621	
Specified Dose Range	16.0 - 45.7 kGy	
SAL	10E-6	
Quantity	260 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1301004B	DK00481250	2025-06-11	25.6
EC__1300947B	DK00481251	2025-06-11	26.4
EC__1299776B	DK00481256	2025-06-17	27.1
EC__1299784B	DK00481252	2025-06-17	30.2
EC__1299818B	DK00481253	2025-06-17	28.1
EC__1299678B	DK00481254	2025-06-17	27.0
EC__1299790B	DK00481255	2025-06-17	26.7
EC__1300963B	DK00481257	2025-06-18	25.5
EC__1301044B	DK00481261	2025-06-18	26.2
EC__1300910B	DK00481258	2025-06-18	26.2
EC__1299685A	DK00481259	2025-06-18	25.7
EC__1299769A	DK00481260	2025-06-18	25.7
EC__1300924B	DK00481262	2025-06-18	26.4

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature