

Certificate of Irradiation

Product	Nunc EasYFlasks 75cm ² w. Filter	Date 2025-06-19
		Page 1
Item Number	156499	
Lot Number	186221	
Specified Dose Range	16.0 - 45.7 kGy	
SAL	10E-6	
Quantity	340 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300961B	DK00475023	2025-06-11	26.3
EC__1300933B	DK00475021	2025-06-11	26.0
EC__1300994B	DK00475022	2025-06-11	26.2
EC__1300377A	DK00475024	2025-06-12	26.1
EC__1300695A	DK00475025	2025-06-12	26.2
EC__1300404B	DK00475026	2025-06-12	28.5
EC__1300320B	DK00475027	2025-06-12	29.6
EC__1300348B	DK00475028	2025-06-12	30.8
EC__1300362B	DK00475029	2025-06-12	31.7
EC__1300376B	DK00475030	2025-06-13	30.6
EC__1300432B	DK00475031	2025-06-13	30.4
EC__1300447A	DK00475032	2025-06-13	30.4
EC__1300428A	DK00475033	2025-06-13	30.9
EC__1301059B	DK00475034	2025-06-18	26.9
EC__1299661B	DK00475036	2025-06-18	21.0
EC__1301016B	DK00475035	2025-06-18	22.5
EC__1299732A	DK00475037	2025-06-18	20.3

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature