
Certificate of Irradiation

Product	Nunc V96 Microwell Plate, Clear, Without Lid	Date 2025-04-20
		Page 1
Item Number	249662	
Lot Number	185378	
Specified Dose Range	21.9 - 47.3 kGy	
SAL	10E-3	
Quantity	480 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1717720A	DK00517595	2025-04-19	26.0
EA__1717704A	DK00517594	2025-04-19	26.6
EA__1718913A	DK00517596	2025-04-19	28.4
EA__1718897A	DK00517597	2025-04-19	27.2
EA__1718925A	DK00517598	2025-04-19	26.3
EA__1717711B	DK00517599	2025-04-19	28.6

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature