

---

## Certificate of Irradiation

<b>Product</b>	Nunc Petri Dish 100 x 15	<b>Date</b> 2025-06-29
		<b>Page</b> 1
<b>Item Number</b>	263991	
<b>Lot Number</b>	186350	
<b>Specified Dose Range</b>	22.5 - 54.6 kGy	
<b>SAL</b>	10E-3	
<b>Quantity</b>	270 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1723914B	DK00478726	2025-06-28	27.6
EA__1723903B	DK00478725	2025-06-28	26.9
EA__1723946B	DK00478724	2025-06-28	27.5
EA__1723943A	DK00478723	2025-06-28	27.0
EA__1723902B	DK00478722	2025-06-28	26.4
EA__1723919B	DK00478721	2025-06-28	26.8
EA__1723915B	DK00478727	2025-06-28	28.3
EA__1724095B	DK00478729	2025-06-28	26.8
EA__1724213A	DK00478728	2025-06-28	26.7

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager

Please note that this is an electronically generated signature