

Certificate of Irradiation

Product	Nunc F96 MicroWell Plates Polystyrene, Clear	Date 2025-08-19
		Page 1
Item Number	161093	
Lot Number	186679	
Specified Dose Range	16.4 - 29.4 kGy	
SAL	10E-3	
Quantity	340 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300097B	DK00517473	2025-08-16	19.3
EC__1299993B	DK00517472	2025-08-16	19.7
EC__1300055B	DK00517474	2025-08-16	19.5
EC__1300027B	DK00517475	2025-08-16	19.1
EC__1300013B	DK00517476	2025-08-16	19.4
EC__1300096A	DK00517477	2025-08-16	19.3
EC__1300142A	DK00517488	2025-08-17	19.1
EC__1300140B	DK00517487	2025-08-17	19.4
EC__1300021B	DK00517486	2025-08-17	19.4
EC__1300061B	DK00517485	2025-08-17	18.6
EC__1300083B	DK00517484	2025-08-17	19.5
EC__1300075B	DK00517483	2025-08-17	18.9
EC__1300013A	DK00517482	2025-08-17	18.9
EC__1300027A	DK00517481	2025-08-17	18.6
EC__1300041A	DK00517480	2025-08-17	18.8
EC__1300103B	DK00517479	2025-08-17	18.9
EC__1296449B	DK00517478	2025-08-17	18.6

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature