
Certificate of Irradiation

Product	12 ml Nunc Test Tube, Sterile	Date	2025-06-29
		Page	1
Item Number	343141		
Lot Number	186168		
Specified Dose Range	21.5 - 38.5 kGy		
SAL	10E-3		
Quantity	80 boxes		

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1724215A	DK00491315	2025-06-28	28.9
EA__1723912B	DK00491316	2025-06-28	28.5

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature