
Certificate of Irradiation

Product	Nunc EasYFlasks 25cm ² w. Filter	Date 2025-06-04
		Page 1
Item Number	156367	
Lot Number	185073	
Specified Dose Range	15.4 - 43.1 kGy	
SAL	10E-6	
Quantity	4 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300892B	DK00475987	2025-06-04	33.5

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature