
Certificate of Irradiation

Product	Nunc EasYFlasks 25cm ² w. Filter	Date 2025-06-27
		Page 1
Item Number	156367	
Lot Number	185313	
Specified Dose Range	15.4 - 43.1 kGy	
SAL	10E-6	
Quantity	122 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1723869A	DK00475994	2025-06-26	20.0
EA__1723901B	DK00475993	2025-06-26	19.0
EA__1723917B	DK00475992	2025-06-26	18.2
EA__1723929B	DK00475991	2025-06-26	18.3
EA__1723916B	DK00475990	2025-06-26	20.1

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature