

Certificate of Irradiation

Product	Nunc F96 MicroWell Polystyrene Plate, White	Date 2025-08-19
		Page 1
Item Number	136101	
Lot Number	186779	
Specified Dose Range	17.1 - 45.9 kGy	
SAL	10E-3	
Quantity	1080 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1299891A	DK00511800	2025-08-15	18.4
EC__1299919A	DK00511801	2025-08-15	18.9
EC__1299905A	DK00511802	2025-08-15	18.8
EC__1299948B	DK00511803	2025-08-15	19.3
EC__1299976B	DK00511804	2025-08-15	18.7
EC__1299878B	DK00511806	2025-08-15	19.3
EC__1299962B	DK00511805	2025-08-15	18.5
EC__1299989B	DK00511814	2025-08-15	19.7
EC__1299851A	DK00511808	2025-08-15	18.5
EC__1299864B	DK00511807	2025-08-15	18.0
EC__1299990B	DK00511809	2025-08-15	18.6
EC__1299849A	DK00511810	2025-08-15	18.9
EC__1299863A	DK00511811	2025-08-15	18.2
EC__1299877A	DK00511812	2025-08-15	18.8
EC__1299835A	DK00511813	2025-08-15	19.0

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature