

Certificate of Irradiation

Product	Nunc 48 Well Multidishes	Date 2025-08-03
		Page 1
Item Number	150687	
Lot Number	185446	
Specified Dose Range	15.8 - 45.0 kGy	
SAL	10E-3	
Quantity	305 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA_1723927B	DK00471576	2025-08-02	19.7
EA_1724166B	DK00471575	2025-08-02	19.0
EA_1723911B	DK00471574	2025-08-02	18.6
EA_1723755B	DK00471580	2025-08-02	21.4
EA_1723737B	DK00471579	2025-08-02	18.6
EA_1724149B	DK00471578	2025-08-02	18.8
EA_1724150B	DK00471577	2025-08-02	19.6

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager
Please note that this is an electronically generated signature