
Certificate of Irradiation

Product	Nunc Centrifuge Tube, 11 ml	Date 2025-03-28
		Page 1
Item Number	347880	
Lot Number	184951	
Specified Dose Range	21.5 - 38.5 kGy	
SAL	10E-6	
Quantity	120 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1723921A	DK00488637	2025-03-27	27.1
EA__1723958B	DK00488636	2025-03-27	26.5
EA__1724120A	DK00488638	2025-03-27	27.3

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature