

## Certificate of Irradiation

<b>Product</b>	Nunc IVF 4-Well Dishes, Non Treated	<b>Date</b> 2025-06-17
		<b>Page</b> 1
<b>Item Number</b>	179830	
<b>Lot Number</b>	185688	
<b>Specified Dose Range</b>	26.5 - 41.7 kGy	
<b>SAL</b>	10E-6	
<b>Quantity</b>	952 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300755B	DK00488801	2025-06-03	31.2
EC__1300895B	DK00488802	2025-06-03	32.4
EC__1300697A	DK00488803	2025-06-03	32.0
EC__1300655A	DK00488804	2025-06-03	31.8
EC__1300749B	DK00488800	2025-06-03	32.8
EC__1300641A	DK00488805	2025-06-03	31.3
EC__1300805B	DK00488807	2025-06-03	31.5
EC__1300777B	DK00488806	2025-06-03	32.1
EC__1300853B	DK00488808	2025-06-03	32.3

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager

Please note that this is an electronically generated signature