

Certificate of Irradiation

Product	Nunc F96 MicroWell Polystyrene Plate, Black	Date 2025-05-25
		Page 1
Item Number	137101	
Lot Number	184958	
Specified Dose Range	17.1 - 45.9 kGy	
Target dose	BATCH XXXXXX	
SAL	10E-3	
Quantity	864 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300040B	DK00511738	2025-05-23	17.9
EC__1300154A	DK00511734	2025-05-23	19.8
EC__1299996A	DK00511735	2025-05-23	18.6
EC__1300024A	DK00511736	2025-05-23	18.9
EC__1300038A	DK00511737	2025-05-23	18.3
EC__1299580B	DK00511739	2025-05-23	18.3
EC__1299566B	DK00511740	2025-05-23	19.4
EC__1300167B	DK00511745	2025-05-24	19.1
EC__1300302B	DK00511744	2025-05-24	19.4
EC__1298748B	DK00511743	2025-05-24	19.3
EC__1300288B	DK00511742	2025-05-24	19.6
EC__1300719A	DK00511741	2025-05-24	19.5

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature