
Certificate of Irradiation

Product	Nunc EasYFlasks, w. Filter	Date 2025-07-09
		Page 1
Item Number	159926	
Lot Number	186718	
Specified Dose Range	15.4 - 43.1 kGy	
SAL	10E-6	
Quantity	51 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1723982A	DK00493208	2025-07-08	17.9
EA__1724027A	DK00493207	2025-07-08	17.9

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack

Senior Quality Assurance Manager

Please note that this is an electronically generated signature