

---

## Certificate of Irradiation

<b>Product</b>	Nunc 1 µl Inoculation Loop, PS, Clear	<b>Date</b> 2025-06-19
		<b>Page</b> 1
<b>Item Number</b>	253287	
<b>Lot Number</b>	185591	
<b>Specified Dose Range</b>	22.4 - 55.5 kGy	
<b>SAL</b>	10E-6	
<b>Quantity</b>	168 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1299819A	DK00491267	2025-06-18	48.4
EC__1299815B	DK00491268	2025-06-18	48.5

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack

Senior Quality Assurance Manager

Please note that this is an electronically generated signature