

---

## Certificate of Irradiation

<b>Product</b>	Deep Well Cap Mat Q07-L243_Nuncatom_PQ108-V02	Date 2025-06-24
		Page 1
<b>Item Number</b>	9503233	
<b>Lot Number</b>	2514610	
<b>Dose target limits</b>	29.5 - 44.2 kGy	
<b>Requested dose</b>	25 - 45 kGy	
<b>Quantity</b>	2 boxes	

Thermo Fisher Scientific confirms that the product supplied has been irradiated according to the agreement with the customer.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1301023A	DK00499010	2025-06-23	32.5

Sub-contract irradiation of customer products is conducted according to the customer's specifications in accordance with ISO11137. "Sterilization of health care products - Radiation". The products have received the requested dose if the min. and max. dose within the dose target limits.

Quality Management System is in compliance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager

Please note that this is an electronically generated signature