
Certificate of Irradiation

Product	Nunc 96-Well Cap Mat, natural, Sterile	Date 2025-08-21
		Page 1
Item Number	276000	
Lot Number	186414	
Specified Dose Range	16.7 - 46.0 kGy	
SAL	10E-3	
Quantity	171 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1299594B	DK00491377	2025-08-20	17.9

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature