

Certificate of Irradiation

Product	Nunc Dish Nunclon 150 x 20	Date 2025-07-09
		Page 1
Item Number	168381	
Lot Number	185627	
Specified Dose Range	16.0 - 45.7 kGy	
SAL	10E-3	
Quantity	503 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1299640B	DK00498190	2025-05-16	27.8
EC__1300104A	DK00498189	2025-05-16	27.3
EC__1301111A	DK00498191	2025-05-16	22.2
EC__1299599A	DK00498193	2025-05-16	21.4
EC__1301097A	DK00498197	2025-05-17	18.6
EC__1299613A	DK00498195	2025-05-17	18.2
EC__1299571A	DK00498194	2025-05-17	19.1
EC__1300426B	DK00498198	2025-05-31	27.4
EC__1299812B	DK00498202	2025-06-17	32.2
EC__1299747B	DK00498201	2025-06-17	32.3
EC__1300932B	DK00498200	2025-06-17	31.4
EC__1301061B	DK00498199	2025-06-17	32.2
EC__1299714B	DK00498203	2025-06-17	32.0
EC__1299700B	DK00498206	2025-06-17	32.5
EA__1724024A	DK00498204	2025-07-09	19.7
EA__1724184A	DK00498205	2025-07-09	19.7
EA__1724169A	DK00498207	2025-07-09	19.4
EA__1724168A	DK00498208	2025-07-09	19.6
EA__1724185A	DK00498209	2025-07-09	20.6

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature