
Certificate of Irradiation

Product	Nunc Dish Nunclon 35 x 10	Date 2025-06-27
		Page 1
Item Number	153066	
Lot Number	185819	
Specified Dose Range	15.8 - 44.3 kGy	
SAL	10E-3	
Quantity	481 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1724223B	DK00397594	2025-06-26	19.2
EA__1724218B	DK00397590	2025-06-26	19.2
EA__1723940A	DK00397589	2025-06-26	19.6
EA__1723888A	DK00397591	2025-06-26	19.5
EA__1723889A	DK00397592	2025-06-26	18.7
EA__1723924A	DK00397593	2025-06-26	19.0

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature