

Certificate of Irradiation

Product	Nunc F96 MicroWell Plate, Clear with Lid, Nunclon	Date 2025-06-06
		Page 1
Item Number	167008	
Lot Number	186282	
Specified Dose Range	17.1 - 45.9 kGy	
SAL	10E-3	
Quantity	360 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300465B	DK00517370	2025-06-05	19.4
EC__1300452B	DK00517369	2025-06-05	20.0
EC__1300468A	DK00517368	2025-06-05	20.6
EC__1300482A	DK00517371	2025-06-05	20.2
EC__1300775A	DK00517372	2025-06-05	19.9

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature