

Certificate of Irradiation

Product	Nunc Dish Nunclon 100 x 15	Date	2025-07-09
		Page	1
Item Number	150350		
Lot Number	186520		
Specified Dose Range	15.2 - 41.8 kGy		
SAL	10E-3		
Quantity	896 boxes		

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1723910A	DK00478731	2025-06-26	17.5
EA__1723942A	DK00478730	2025-06-26	18.0
EA__1723926A	DK00478732	2025-06-26	16.9
EA__1724184B	DK00478733	2025-07-08	18.7
EA__1724142B	DK00478734	2025-07-08	17.9
EA__1724143B	DK00478735	2025-07-08	18.5
EA__1724070B	DK00478736	2025-07-08	18.6
EA__1724071B	DK00478737	2025-07-08	18.9
EA__1724053B	DK00478738	2025-07-08	18.6
EA__1724055B	DK00478739	2025-07-08	18.6
EA__1724126B	DK00478740	2025-07-08	19.7
EA__1723995A	DK00478743	2025-07-08	18.2
EA__1723997A	DK00478741	2025-07-08	19.0
EA__1723979A	DK00478742	2025-07-08	18.6

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature