

Certificate of Irradiation

Product	Nunc IVF Petri Dishes, Non Treated	Date 2025-06-04
		Page 1
Item Number	150270	
Lot Number	185813	
Specified Dose Range	25.3 - 45.2 kGy	
SAL	10E-6	
Quantity	1280 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300531A	DK00448181	2025-06-02	33.5
EC__1300471A	DK00448182	2025-06-02	30.5
EC__1300545A	DK00448183	2025-06-02	30.3
EC__1300539A	DK00448184	2025-06-02	30.8
EC__1300553A	DK00448185	2025-06-02	30.3
EC__1300474A	DK00448186	2025-06-02	31.1
EC__1300506B	DK00448187	2025-06-02	30.6
EC__1300593B	DK00448188	2025-06-02	30.7
EC__1300344A	DK00448190	2025-06-02	30.9
EC__1300559A	DK00448191	2025-06-02	31.1
EC__1300573A	DK00448192	2025-06-02	30.9
EC__1300721A	DK00448193	2025-06-02	30.8
EC__1300526A	DK00448194	2025-06-02	31.0
EC__1300530A	DK00448195	2025-06-02	30.8
EC__1300810B	DK00448196	2025-06-02	31.4
EC__1300796B	DK00448197	2025-06-02	31.0
EC__1300838B	DK00448199	2025-06-02	31.8
EC__1300329B	DK00448198	2025-06-02	31.2
EC__1300832B	DK00448200	2025-06-02	31.7
EC__1300795B	DK00448201	2025-06-02	31.2
EC__1300790B	DK00448204	2025-06-02	31.7
EC__1300804B	DK00448203	2025-06-02	31.2
EC__1300754B	DK00448202	2025-06-02	30.9
EC__1300402A	DK00448205	2025-06-02	32.2
EC__1300783B	DK00448208	2025-06-03	30.9
EC__1300797B	DK00448207	2025-06-03	30.4
EC__1300811B	DK00448206	2025-06-03	30.4

Certificate of Irradiation

Product	Nunc IVF Petri Dishes, Non Treated	Date 2025-06-04
		Page 2
Item Number	150270	
Lot Number	185813	
Specified Dose Range	25.3 - 45.2 kGy	
SAL	10E-6	
Quantity	1280 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
---------------------	----------	---------------------	----------

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature