

Certificate of Irradiation

Product	Nunc IVF Petri Dishes, Non Treated	Date 2025-06-27
		Page 1
Item Number	150270	
Lot Number	185877	
Specified Dose Range	25.3 - 45.2 kGy	
SAL	10E-6	
Quantity	1279 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300473B	DK00397577	2025-06-25	29.9
EC__1300488B	DK00448213	2025-06-25	31.4
EC__1299645A	DK00448212	2025-06-25	31.1
EC__1301105A	DK00448211	2025-06-25	31.3
EC__1300459B	DK00448210	2025-06-25	30.4
EC__1300502B	DK00448214	2025-06-25	29.5
EC__1300516B	DK00448215	2025-06-25	29.2
EC__1299567A	DK00448216	2025-06-25	30.6
EC__1299744B	DK00448217	2025-06-25	30.2
EC__1299644A	DK00448218	2025-06-25	30.1
EC__1301077A	DK00448219	2025-06-25	30.6
EC__1299579A	DK00448220	2025-06-25	31.0
EA__1723932A	DK00448221	2025-06-25	32.1
EA__1723919A	DK00448222	2025-06-25	31.7
EA__1724208B	DK00448223	2025-06-25	31.0
EA__1723908A	DK00448224	2025-06-25	31.7
EA__1723899A	DK00448225	2025-06-25	31.2
EA__1723915A	DK00448226	2025-06-25	31.1
EA__1723947A	DK00448227	2025-06-25	31.5
EA__1723931A	DK00448228	2025-06-25	32.3
EA__1723865A	DK00448229	2025-06-25	32.2
EA__1723890A	DK00448230	2025-06-25	31.6
EA__1724217B	DK00448231	2025-06-25	31.5
EA__1723903A	DK00448232	2025-06-25	31.4
EA__1724221B	DK00448233	2025-06-25	31.6
EA__1723927A	DK00448234	2025-06-25	34.5
EA__1724216B	DK00448209	2025-06-26	32.5

Certificate of Irradiation

Product	Nunc IVF Petri Dishes, Non Treated	Date 2025-06-27
		Page 2
Item Number	150270	
Lot Number	185877	
Specified Dose Range	25.3 - 45.2 kGy	
SAL	10E-6	
Quantity	1279 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
---------------------	----------	---------------------	----------

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature