

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or distributed according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.



Alan E. Hatch
 Sr. Quality Manager
 11/07/2025

The following information represents Product Certification for: Item#: **3640-0025**

Description: **BURET SELF-ZEROING ACRL;25ML**

Lot#: **1440657**

Manufactured: **10/25/2025**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-1206-85	GROMMET,WASH/BTL	COMPONENT PART				
8-0049-01	RESIN,LDPE,INJ	LOW-DENSITY POLYETHYLENE	9522	PASSED	PASSED	177.1520(a)(2)(i)
1-1811-10	CLOS,53/415,PP,NAT,NALGE	COMPONENT PART				
8-0071-06	Resin,PP,Inj	POLYPROPYLENE, INJECTION	9988	PASSED	PASSED	177.1520(a)(1)(i), (c)1.1a,177.1520(b), (use conditionsA-H)
1-1206-98	BTL,500ML,RND,W/M,LDPE	COMPONENT PART				
8-0049-31	RESIN,LDPE,IBM,EBM,EXT	LOW-DENSITY POLYETHYLENE	1572	PASSED	PASSED	176.170(c), 177.1520(c)2.2,use conditions B-H
1-3432-95	FLNG,SIDE ARM,10&25ML,BURET	COMPONENT PART				
8-0092-02	RESIN,ACR,TERPOLYMER,INJ	POLYMETHYL METHACRYLATE	BB-MF-2138	PASSED	PASSED	177.1010, 176.170
1-3435-64	HDL,STPCOK,PLUG,PP,BLK	COMPONENT PART				
8-0071-15P	RESIN,PP,BLK,INJ	COMPONENT PART				
8-0071-06	Resin,PP,Inj	POLYPROPYLENE, INJECTION	9988	PASSED	PASSED	177.1520(a)(1)(i), (c)1.1a,177.1520(b), (use conditionsA-H)
8-0099-43	COLOR,BLK,MULTI	COLORANT, BLACK	N/A	N/A	N/A	177.1350, 1520, 1560, 2600,178.3297, 181.28
1-3435-70	HSG,STPCOK,25ML,BURET,PMP	COMPONENT PART				
8-0022-01	RESIN,PMP,INJ	POLYMETHYLPENTENE	3621	PASSED	PASSED	177.1520(a)(3)(ii)
1-3435-84	TIP,BURET,PMP,10 & 25ML	COMPONENT PART				
8-0022-01	RESIN,PMP,INJ	POLYMETHYLPENTENE	3621	PASSED	PASSED	177.1520(a)(3)(ii)
1-3705-01	COV,BURET	COMPONENT PART				
8-0049-01	RESIN,LDPE,INJ	LOW-DENSITY POLYETHYLENE	9522	PASSED	PASSED	177.1520(a)(2)(i)

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or distributed according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.



Alan E. Hatch
Sr. Quality Manager
11/07/2025

The following information represents Product Certification for: Item#: **3640-0025**

Description: **BURET SELF-ZEROING ACRL;25ML**

Lot#: **1440657**

Manufactured: **10/25/2025**

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Burets have been individually calibrated to meet the accuracy requirements of ASTM E287.