

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625 Product Certificate Thermo Scientific Nalgene and Nunc Products

Alm E Hatel

Alan E. Hatch Sr. Quality Manager 06/26/2025

06/20/2025

Manufactured:

or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.

management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694)

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or

distributed according to the requirements of product and quality specifications as maintained in our quality

The following information represents Product Certification for: Item#: 342821-0116

Description: CLOS, MPV, LOW PRO, BLUE, PPCO; 11MM Lot#: 1431400

USP Class VI Description Common Name DMF# Cytotoxicity FDA Compliance - 21 CFR Part Number COMPONENT PART 1-3828-62P LPCLOS MPV PPCO BLU W/O TETHER 8-0028-15P RESIN, PPCO, RAD STAB, BLUE, INJ COMPONENT PART 8-0028-04 RESIN, PPCO, RAD STER, INJ POLYPROPYLENE COPOLYMER 7478 PASSED PASSED 177.1520 (a)(3)(i) & (c)3.1(a)except for cooking, (useconditions C-H) 8-0099-74 COLOR, BLUE, MULTI COLORANT, BLUE, MULTI-RESIN N/A PASSED PASSED 177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210

Use Before: 06/20/2030

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Product was Gamma Irradiation Sterilized. Product was dosimetric released per ANSI/AAMI/ISO 11137 guidelines. Product was determined to be non-pyrogenic at a level < 0.5 EU/ml per USP < 85 > .

Product is certified to be free of detectable RNase/DNase contamination. This test is performed using the nuclease assay method with a detection limit of 8 x 10<sup>-7</sup> Kunitz unit/ul for DNase and 1.9 x 10<sup>-10</sup> Kunitz unit/ul for RNase.