

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or distributed according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.



Alan E. Hatch  
 Sr. Quality Manager  
 06/19/2025

The following information represents Product Certification for: Item#: **264575**

Description: **384 WELL PP PLATE WHITE NONST**

Lot#: **1429919**

Manufactured: **06/10/2025**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
009326A00P	LID,NUNC,384 WELL PLATE,PS,CLR	COMPONENT PART				
14149MR	RESIN,POLYSTYRENE	NATURAL, POLYSTYRENE, INJ.	18492	PASSED	PASSED	177.1640
009402B01P	384 PLATE, NUNC, PP, WHITE	COMPONENT PART				
14285MRP	BLEND, RESIN, WHITE, PP	COMPONENT PART				
14117MR	RESIN,PP,HIMONT PF-511	PP, HOMO, NATURAL, INJ.	26106	PASSED	PASSED	177.1520(a)(1)(i) and (c)1.1a, 177.1520(b), use conditions A-H
14284MR	COLORANT,WHITE,PP	COLORANT, WHITE, PP, INJ.	N/A	N/A	N/A	177.1520 3.1a, not for articles used for holding food during cooking, 178.2010, 3297

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.