

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or distributed according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.

The following information represents Product Certification for: Item#: **1600-0032**

Description: **BTL N/M FEP;32OZ,1000ML**

Lot#: **1420285**

*Alan E Hatch*

Alan E. Hatch  
Sr. Quality Manager  
02/05/2025

Manufactured: **12/31/2024**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-1843-84	CLOS,38/415,ETFE,NAT,NALGE	COMPONENT PART				
8-0008-01	RESIN,ETFE,INJ	ETHYLENE-TETRAFLUOROETHYLENE	N/A	PASSED	PASSED	177.1550 at temperaures up to 121C
1-0400-89PP	BTL,1000ML,RND,N/M,FEP	COMPONENT PART				
8-0015-36	RESIN,FEP,EBM	RESIN,FEP,EBM	N/A	PASSED	PASSED	177.1550

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.