

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or distributed according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.



 Alan E. Hatch
 Sr. Quality Manager
 11/11/2025

 The following information represents Product Certification for: Item#: **3743RED**

 Description: **Tube, ScrewTop, 500 uL 2D V-Bt**

 Lot#: **1441725**

 Manufactured: **11/06/2025**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-0024-79	CLSR,2SHOT,THREADED SCREW TUBE,RED	COMPONENT PART				
139-0048-00	Resin, RP PPCO, Inj	POLYPROPYLENE COPOLYMER	N/A	PASSED	PASSED	177.1520(a)(3)(i) and 177.1520(b) (use conditions C through H)
8-0005-25	RESIN,TPV,NAT,75 DURO	THERMOPLASTIC RUBBER	12719	PASSED	PASSED	N/A
8-0097-35	COLOR,RED MATRIX	COLORANT, RED MATRIX	N/A	PASSED	PASSED	177.1520, 178.2010, 178.3297, 178.3710
23464	Tube, screw, 500uL, V-Bottom	COMPONENT PART				
139-0080-00	Resin, P5 PPCO, Inj	POLYPROPYLENE COPOLYMER	9988	PASSED	PASSED	177.1520(a)(3)(i) and 177.1520(c)3.1a

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.