

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

Product Certificate

Thermo Scientific Nalgene and Nunc Products

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or distributed according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.

Alan E. Hatch Sr. Quality Manager 06/18/2025

The following information represents Product Certification for: Item#: 249949

Description: V96 POLYPROP PLATE WHITE NS Lot#: 1429917 Manufactured: 05/28/2025

| Part Number | Description | Common Name | DMF# | Cytotoxicity | USP Class VI | FDA Compliance - 21 CFR |
|-------------|--------------------------------|----------------------------|-------|--------------|--------------|--|
| 009326A00P | LID,NUNC,384 WELL PLATE,PS,CLR | COMPONENT PART | | | | |
| 14149MR | RESIN, POLYSTYRENE | NATURAL, POLYSTYRENE, INJ. | 18492 | PASSED | PASSED | 177.1640 |
| 009367B04P | V96 PLATE, NUNC, PP, WHITE | COMPONENT PART | | | | |
| 14285MRP | BLEND, RESIN, WHITE, PP | COMPONENT PART | | | | |
| 14117MR | RESIN, PP, HIMONT PF-511 | PP, HOMO, NATURAL, INJ. | 26106 | PASSED | PASSED | 177.1520(a)(1)(i) and (c)1.1a, 177.1520(b), use conditions A-H |
| 14284MR | COLORANT,WHITE,PP | COLORANT, WHITE, PP, INJ. | N/A | N/A | N/A | 177.1520 3.1a, not forarticles used for holding foodduring cooking, 178.2010, 3297 |

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Product produced after Lot 570055 is certified to be free of detectable RNase/DNase contamination. This test is performed using the nuclease assay method with a detection limit of 8 x 10^ -7 Kunitz unit/ul for DNase and 1.9 x 10^ -10 Kunitz unit/ul for RNase.

Product certified to be free of PCR inhibitors and detectable Human DNA contamination using a real time PCR assay method that certifies the mean quantity of DNA for test samples and negative controls with a detection limit of 0.8 pg / μ L.