

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625 Product Certificate Thermo Scientific Nalgene and Nunc Products

Alm E Hatel

Alan E. Hatch Sr. Quality Manager 06/06/2025

The following information represents Product Certification for: Item#: 382019-2000

or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or

management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694)

distributed according to the requirements of product and quality specifications as maintained in our quality

Description: SQ BTL STR Clean PETG, 2000mL

Lot#: 1429675

Use Before: 05/28/2030

Manufactured: 05/28/2025

| Part Number | Description | Common Name | DMF# | Cytotoxicity | USP Class VI | FDA Compliance - 21 CFR |
|-------------|-------------------------------|---------------------------|------|--------------|--------------|-----------------------------|
| 1-0638-91P | BTL,2L,SQ,N/M,PETG,MEDIA,2cav | COMPONENT PART | | | | |
| 8-0001-32 | RESIN, PETG, IBM, EBM, INJ | COPOLYESTER, PETG NATURAL | 9987 | PASSED | PASSED | 177.1315(b)(1) and 174.5 |
| 1-1822-21 | CLOS,53B,HDPE,NAT | COMPONENT PART | | | | |
| 8-0042-01 | RESIN, HDPE, INJ | HIGH-DENSITY POLYETHYLENE | 1646 | PASSED | PASSED | 176.170(c), 177.1520(c)3.2a |

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Product was Gamma Irradiation Sterilized. Product was dosimetric released per ANSI/AAMI/ISO 11137 guidelines. Product was determined to be non-pyrogenic at a level < 0.5 EU/ml per USP < 85 > .

Certified clean containers are manufactured in an ISO 14644-1 Class 7 cleanroom. Liquid particle count levels are lot certified in compliance with USP < 788 > , EP 2.9.19, and JP 14th Edition Part 1 - Section 24.

The materials used in the manufacture of these resins are free of animal derived material.

ANIMAL DERIVED MATERIALS (BSE/TSE) - All resins and colorants used at the manufacturing site comply with the latest revision of EMA/410/01 section 6.4.