

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

## **Product Certificate**

Thermo Scientific Nalgene and Nunc Products

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or distributed according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.

Alan E. Hatch Sr. Quality Manager 10/21/2025

Alm E Hatel

The following information represents Product Certification for: Item#: **3744GRE-BR** 

Description: Mtx 0.5ml, 2D,V-Btm w/Green Cap, BC Rack Lot#: 1440337

Manufactured: 10/17/2025

| Part Number | Description                          | Common Name             | DMF#  | Cytotoxicity | USP Class VI | FDA Compliance - 21 CFR  |
|-------------|--------------------------------------|-------------------------|-------|--------------|--------------|--|
| 1-0024-82   | CLSR,2SHOT,THREADED SCREW TUBE,GREEN | COMPONENT PART          |       |              |              |  |
| 139-0048-00 | Resin, RP PPCO, Inj                  | POLYPROPYLENE COPOLYMER | N/A   | PASSED       | PASSED       | 177.1520(a)(3)(i) and 177.1520(b) (use conditions C through H) |
| 8-0005-25   | RESIN,TPV,NAT,75 DURO                | THERMOPLASTIC RUBBER    | 12719 | PASSED       | PASSED       | N/A  |
| 8-0097-38   | COLOR, GREEN MATRIX                  | COLORANT, GREEN MATRIX  | N/A   | N/A          | N/A          | 177.1520, 178.2010, 178.3297, 178.3710                         |
| 23464       | Tube, screw, 500uL, V-Bottom         | COMPONENT PART          |       |              |              |  |
| 139-0080-00 | Resin, P5 PPCO, Inj                  | POLYPROPYLENE COPOLYMER | 9988  | PASSED       | PASSED       | 177.1520(a)(3)(i) and<br>177.1520(c)3.1a                       |

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.