

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or distributed according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.



Alan E. Hatch
 Sr. Quality Manager
 11/15/2025

The following information represents Product Certification for: Item#: **374222-BR**

Description: **Nunc Cryobank 5.0mL, Non-2D,w/Cap, BC Rk** Lot#: **1440716**

Use Before: **10/25/2030**

Manufactured: **10/25/2025**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-0025-71	VIAL,CRYOBANK,5.0mL,W/BC HEAD	COMPONENT PART				
8-0071-22	Resin,PP,Black,INJ,Cryobank	COMPONENT PART				
14117MR	RESIN,PP,HIMONT PF-511	PP, HOMO, NATURAL, INJ.	26106	PASSED	PASSED	177.1520(a)(1)(i) and (c)1.1a, 177.1520(b), use conditions A-H
14288MR	COLORANT,BLACK,PP	COLORANT, BLACK, PP, INJ.	N/A	N/A	N/A	N/A
8-0028-04	RESIN,PPCO,RAD STER,INJ	POLYPROPYLENE COPOLYMER	7478	PASSED	PASSED	177.1520 (a)(3)(i) & (c)3.1(a)except for cooking, (useconditions C-H)
1-0025-73	CAP,CRYOVIAL 2/5mL NON-PIERCEABLE	COMPONENT PART				
8-0028-04	RESIN,PPCO,RAD STER,INJ	POLYPROPYLENE COPOLYMER	7478	PASSED	PASSED	177.1520 (a)(3)(i) & (c)3.1(a)except for cooking, (useconditions C-H)
8-0005-22	RESIN,TPV,NAT,45 DURO SANTOPRENE	THERMOPLASTIC RUBBER	12719	PASSED	PASSED	N/A

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Product was E-Beamed Sterilized. Product was dosimetric released per ANSI/AAMI/ISO 11137 guidelines. Product was determined to be non-pyrogenic at a level < 0.5 EU/ml per USP < 85 > .

Product is certified to be free of detectable RNase/DNase contamination. This test is performed using the nuclease assay method with a detection limit of 8 x 10⁻⁷ Kunitz unit/ul for DNase and 1.9 x 10⁻¹⁰ Kunitz unit/ul for RNase.